

Interproximal Caries Detection Ability of Dental Practitioners Using Direct Digital System and Conventional Films

Original Article

Najmeh Movahhedian¹, Leila Khojastepour¹, Fahimeh Mortazavi¹, Mohammad Sadegh Birjandi²

¹Assistant Professor, Department of Oral and Maxillofacial Radiology, Faculty of Dentistry, Shiraz University of Medical Science, Shiraz, Iran

²Dental Practitioner

Received: Mar 11, 2015

Accepted: Apr 9, 2015

Corresponding Author:

Leila Khojastepour

Address:

No. 232, between Molasadra and Felestin Avenue, Moadel Avenue, Shiraz, Iran

Telephone: +989173158331

E-mail: khojastl@sums.ac.ir

Abstract

Introduction:

The study aimed to compare the accuracy of proximal caries detection by dental practitioners using two different systems, direct digital system and conventional films, under routine environmental conditions.

Materials and methods:

Eleven extracted human posterior teeth were mounted on wax frames and were exposed using the direct digital system and conventional films. Seventy-four dental practitioners evaluated these images for proximal caries. No environmental conditions were dictated. The dental practitioners' evaluations were compared with histopathological sections that were used as the gold standard.

Results:

There were no statistically significant differences in the dental practitioners' ability to detect caries using the direct digital and conventional systems ($P=0.548$). No significant differences were found between groups based on the level of the dental practitioners' work experience ($P=0.167$).

Conclusion:

Dental practitioners work well with both direct digital and conventional systems in the diagnosis of proximal caries, although the routine environmental condition is not the ideal one. Further, this ability is so basic that it is not influenced by the level of the dental practitioner's experience. Therefore, we recommend that digital systems be considered for use in dental offices.

Key words:

•Dental Caries •Digital Dental Radiography
•Dentists

Introduction

Dental caries is one of the most common diseases worldwide, and caries detection is of basic importance for every dental practitioner. Although clinical examination is essential for detecting dental caries, it has certain shortcomings even when conducted under ideal conditions, particularly in proximal areas of teeth with close interproximal contacts that lead to false negative diagnoses. Thus, the radiograph is a necessary supplement to clinical examination.^(1,2)

For many years, intraoral radiographs have been taken using conventional films; however, nowadays, digital receptors are increasing in popularity. Digital receptors could be classified as solid-state sensors (CCD/CMOS) or phosphor plates (PSP). The use of direct digital systems based on CCD or CMOS receptors is far more common than phosphor plate-based digital systems because they provide real time images with more reasonable installation costs.

Under optimal conditions, digital intraoral systems enable less patient exposure, eliminate wet processing, save time, and facilitate image communication in comparison with conventional techniques.^(1,3,4,5) Apart from the diagnostic ability of each system, viewing conditions would also affect the ability to detect subtle or even more extensive changes.⁽⁶⁾ Unfortunately, this aspect has been overlooked by dental practitioners owing to either paucity of time or lack of knowledge.

At the time of introduction, the accuracy of the available digital systems was compared with that of conventional films in a number of *in vitro* and *in vivo* studies. However, the results of these studies were based on controlled viewing conditions and educated observers.^(7,8)

The present study attempts to elucidate how the caries detection ability of dental practitioners could be influenced by two different intraoral imaging systems (conventional films and CCD) under routine environmental conditions and determine whether this ability is influenced by the level of the dental practitioners' work experience.

Materials and Methods

In this cross-sectional study, 74 dental practitioners evaluated two sets of radiographs obtained from 11 extracted human posterior teeth

separately: 1 set on the conventional radiographic film and the other on the computer screen, obtained using the CCD receptor. The exercise was repeated after an interval of 2 weeks. No special viewing conditions were dictated. The observers were asked to view both sets of radiographs under the same environmental conditions as their everyday practice and subsequently fill a questionnaire designed for reporting the caries status of each interproximal surface on the radiograph. Teeth selection criteria

A large number of extracted human posterior teeth were evaluated radiographically by 2 oral and maxillofacial radiologists to select 11 teeth (with agreement) that could be categorized under the following 4 groups, according to their interproximal caries status:

1-Sound

2-Confined to DEJ

3-Confined to outer half of the dentinal width

4-Exceeding the outer half of the dentinal width

These teeth were also assessed histologically to determine the actual depth of the caries, which was used as the gold standard. Histological sections that were parallel to the longitudinal axis of the crown were obtained. A dental pathologist examined all the sections under a stereomicroscope ($\times 10$) in order to evaluate the caries status and determine the exact depth of the existing caries. Cases with discrepancies between radiographic diagnosis and histological evaluation were not included in the study.

Radiographs

All the teeth were mounted on 3 wax blocks with the proximal surfaces in contact with each other. The blocks were then adjusted and attached to the bite plate of a Dentsply XCP film holding system (Rinn Corp., Elgin, IL, USA) to standardize the focal-to-object distance (FOD) and focal-to-film distance (FFD) in all the radiographic examinations including conventional and digital projections.

For conventional projections, we used E-speed Kodak film (Eastman Kodak, Rochester, NY, USA) and for digital projections, we used Dr. Suni direct digital intraoral CCD sensor (Suni Medical Imaging, San Jose, CA, USA).

The x-ray unit that was used for all projections was Planmeca IntraX-ray machine (Planmeca Oy, Helsinki, Finland). Exposure factors including kVp, mA, and time were determined by con-

ducting a pilot study to obtain an acceptable optical density. For both systems, the teeth blocks were exposed under different technical conditions and two oral & maxillofacial radiologists decided on the radiograph that had better optical density for the diagnosis of carious lesions. In the case of absence of any difference, the lower exposure factor was chosen.

For processing conventional films, we used an automatic processor (Hope; Dental Max, USA) using fresh Champion processing chemicals (Tehran, Iran).

Statistical Analysis

The Wilcoxon signed rank test was used to compare the diagnostic abilities of dental practitioners in detecting dental caries using the two methods. The Kruskal–Wallis H test was employed to compare the diagnostic abilities of groups with different levels of work experience. The significance level was set at 0.05.

The statistical software SPSS v.15.0 (SPSS Inc., Chicago, IL, USA) was used for data analysis. ($p < 0.05$).

The MDW measurements, ratios and combina

tion charts of the upper and lower central incisors obtained from the Vitapan, Ideal Makoo, and Ivoclar product catalogs, are illustrated in Tables 1, 2, and 3, respectively.

Results

The true proximal surface status of the studied teeth according to histological sections included 12 sound surfaces (S), 2 surfaces in which dental caries was restricted to enamel up to the dentin-enamel junction (DEJ), 2 surfaces in which dental caries involved less than half of the dentin thickness ($<1/2$ Dentin), and 5 surfaces in which dental caries extended to half or more than half of the dentin thickness ($\geq 1/2$ Dentin).

Table 1 shows the median, mean, and standard deviation values of correct diagnosis of proximal surface status using both digital and conventional forms. There was no statistically significant difference in the correct diagnosis of proximal surface status between (sound, dental caries up to DEJ, Dental caries up to half of dentin and dental caries at or more than half of dentine) digital and conventional intraoral radiographs ($P = 0.54$).

Table 1: Correct diagnosis of proximal surface status using digital and conventional forms

Method	Conventional Median Mean±SD	Digital Median Mean±SD	P value
Sound	11.5 11±1.49	12 11.04±1.76	0.560
DEJ	1.00 0.68±0.68	1.00 0.84±0.64	0.071
<1/2 Dentin	0.50 0.73±0.86	1.00 0.77±0.87	0.714
>1/2 Dentin	1.00 1.20±0.76	1.00 1.15±0.70	0.498
Total	15.00 14.34±2.54	15.00 14.50±2.54	0.548

Table 2 summarizes the median, mean, and standard deviation values of correct diagnosis of proximal surface status using digital and conventional intraoral radiographs by 3 groups of dental practitioners with different levels of experience (≤ 5 years, 5–10 years, and ≥ 10 years). There was no statistically significant difference

in the correct diagnosis of proximal surface status between the 3 groups ($P = 0.2$ for conventional films; $P = 0.16$ for digital systems).

Table 2. Correct diagnosis of proximal surface status by three groups of dentists with different work experience using digital and conventional forms

Experience		≤ 5 years	5–10 years	≥ 10 years	P value
Caries group		Median Mean ±SD	Median Mean ±SD	Median Mean ±SD	
Sound	Conventional	11.00 10.63±1.79	12.00 11.13±1.58	12.00 11.22±1.05	0.356
	Digital	12.00 11.04±1.63	12.00 11.04±1.89	12.00 11.04±1.83	0.994
DEJ	Conventional	1.00 0.63±0.65	1.00 0.78±0.74	1.00 0.63±0.69	0.702
	Digital	1.00 0.83±0.64	1.00 1.00±0.60	1.00 0.70±0.67	0.244
<1/2 Dentin	Conventional	0.00 0.46±0.72	1.00 0.83±0.94	1.00 0.89±0.89	0.148
	Digital	0.00 0.71±0.91	1.00 0.83±0.94	1.00 0.78±0.80	0.848
>1/2 Dentin	Conventional	1.00 1.17±0.76	1.00 1.30±0.70	1.00 1.15±0.82	0.777
	Digital	1.00 0.96±0.69	1.00 1.17±0.65	1.00 1.30±0.72	0.207
Total	Conventional	14.00 13.63±2.55	15.00 14.91±2.64	15.00 14.48±2.39	0.227
	Digital	14.00 14.17±2.37	15.00 14.83±2.95	15.00 14.52±2.36	0.167

Discussion

Various studies have compared the accuracy of different direct digital sensors with conventional films.^(8,9,10,11,12) All of these studies were conducted under ideal environmental conditions: conventional films were viewed on a view box under subdued environmental light and digital images were viewed on a monitor of specific size and resolution under subdued environmental light. The observers were well trained and instructed for calibration purposes.^(7,8) Most of these studies concluded that the diagnostic accuracy of direct digital and conventional images were comparable.^(4,7,8, 13–16)

However, in everyday dental practice, dental practitioners usually view and interpret radiographs in conditions that are different from those used in experiments for various reasons, among which are paucity of space and time, and this can affect the result.

In agreement with the above statement is the study by Haak et al.⁽¹⁷⁾ They concluded that smaller gray scale differences could be detected on LCD monitors compared with CRT monitors. In another study, Hellen-Halme et al.⁶ found that multiple environmental factors could affect the diagnostic accuracy of digital radiography. They

mentioned ambient light, screen reflection, and clean monitors among those factors. They concluded that although brightness and contrast settings could affect the diagnosis, ambient light remained a more important factor. Further, Li et al.⁽¹⁸⁾ demonstrated that the length of an endodontic file was more accurately measurable under subdued lighting. However, a questionnaire study showed that 40% of general dental practitioners do not use any quality control for their digital images.⁽¹⁹⁾ In our study, dental practitioners viewed both conventional and digital images under routine environmental conditions and no quality control conditions were dictated. Our results showed that dental practitioners diagnosed interproximal caries in direct digital images with the same accuracy as with conventional E-speed film, and experience (evaluated in terms of years of work experience) did not have a significant effect on this result. In other words, experienced dental practitioners did not show significant differences in their ability to detect interproximal caries as compared with inexperienced ones. This finding is in contrast with a study by Upriched et al.⁽⁹⁾ that showed that with increasing experience, accuracy in detection of proximal caries in mixed dentition using digital systems

increased. However, they also concluded that over time, as digital systems become more popular and dental practitioners get familiar with their use, the accuracy of diagnosis using direct digital images could become comparable with that of conventional films. This seems to be the case with our study conducted 14 years after the mentioned study. During this period, dental practitioners got accustomed to this modality and its quality control options, and this improved their diagnosis, as expected by Upriched et al.⁽⁹⁾ Wenzel⁽²⁰⁾ reviewed 9 questionnaire studies on the use of digital radiography by dental practitioners. In his study, most general dental practitioners preferred using digital systems for different reasons, most common among which were avoidance of chemicals, reduced patient exposure, lesser time, and remote communication, although in some of the reviewed studies, more errors and retakes took place while using digital systems.

Conclusion

Since dental practitioners perform well and comparably using both direct digital and conventional images we recommend that the use of digital

systems in dental offices be considered for different reasons, most important amongst which are avoidance of chemicals, reduced patient exposure, lesser time, and remote communication.

Acknowledgment

This manuscript has been extracted from Ms. Fahimeh Mortazavi's DDS thesis that was conducted under the supervision of Dr. Najmeh Movahhedian and advisory of Dr. Leila Khojastepour. The study was approved, registered with ID 8593057, and supported by the International Branch of Shiraz University of Medical Sciences. The authors wish to thank Dr.Mehradad Vosogh, faculty at Center for Research Improvement, Shiraz School of Dentistry, for the statistical analysis. We also wish to express our gratitude to the Implant Research Center of Shiraz Dental School for providing us the direct digital system.

References

- 1.Pontual AA, Melo DP, Almeida SM, Boscolo FN, HaiterNeto F. Comparison of digital systems and conventional dental film for the detection of approximal enamel caries. *DentomaxillofacRadiol*. 2010;39(7):431-36.
- 2.White SC, Pharoah MJ. *Oral radiology principle and interpretation*.6th ed. Missouri:Mosby.2009:273-5.
- 3.Pretty IA. Caries detection and diagnosis: novel technologies. *J Dent* 2006;34(10):727-39.
- 4.Jacobson JH, Hansen B, Wenzel A, Hintze H. Relationship between histological and radiographic caries lesion depth measured in images from four digital radiography systems. *Caries Res* 2004;38(1):34-8.
- 5.Pai SS, Zimmerman JL. Digital radiographic imaging in dental practice. *Dent Today* 2002 ;21(6):56-61.
- 6.Hellen-Halme K, Nilsson M, PeterssonA. Digital radiography in general dental practice: a field study.2007 ;36(5):249-55.
- 7.Castro VM, Katz JO,HardamPK,GlarosAG,SpencerP.In vitro comparison of conventional film and direct digital imaging in detection of approximal caries. *DentomaxillofacRadiol*2007;36(3):138-42.
- 8.Abreu Junior M,TyndallDA,PlatinE,LudlowJB,PhillipsC.Two- and three-dimensional imaging modalities for the detection of caries. A comparison between film, digital radiography and tuned aperture computed tomography(TACT). *DentomaxillofacRadiol* 1999;28(3):152-7.
- 9.Upriched Kristen K , Potter BJ, Russell CM,Schafer TE, Adair S, Weller RN. Comparison of direct digital and conventional radiography for the detection of proximal surface caries in the mixed dentition. *Pediatr Dent* 2000 ;22(1):9-15.
- 10.Syriopoulos K, Sanderink GCH, VeldersXL,van der Stelt PF. Radiographic detection of approximal caries: a comparison of dental films and digital imaging systems.*DentomaxillofacRadiol* 2000; 29(5):312-8.
- 11.Tyndall DA, Ludlow JB,PlatinE,Nair M.A comparison of Kodak Ektaspeed Plus film and the SiemansSidexis digital imaging system for caries detection using receiver operating characteristic analysis. *Oral Surg Oral Med Oral Pathol Oral RadiolEndod* 1998;85(1):113-8.
- 12.Price C, Ergul N.A comparison of a film-based and a direct digital dental radiographic system using a proximal caries model. *DentomaxillofacRadiol* 1997;26(1):45-52.
- 13.Pine CM, ten Bosch JJ.Dynamics of and diagnostic methods for detecting small carious lesions. *Caries Res*

Interproximal Caries Detection Ability of Dental Practitioners

1996;30(6):381-8.

14. Yang J, Dutra V. Utility of radiology, laser fluorescence, and transillumination. *Dent Clin North Am* 2005;49(4):739-52.

15. Peers A, Hill FJ, Mitropoulos CM, Holloway PJ. Validity and reproducibility of clinical examination, fiber-optic transillumination, and bite-wing radiology for the diagnosis of small approximal carious lesions: an in vitro study. *Caries Res* 1993;27(4):307-11.

16. Wenzel A. Digital imaging for dental caries. *Dent Clin North Am* 2000;44(2):319-38.

17. Haak R, Wicht MJ, Nowak G, Hellmich M, Nowak G, Noack MJ. Influence of room lighting on gray-scale perception with a CRT and a TFT monitor display. *Dentomaxillofac Radiol* 2002;31(3):193-7.

18. Li G, Sanderink GCH, Welander U, McDavid WD, Nasstrom K. Evaluation of endodontic files in digital radiographs before and after employing three image processing algorithms. *Dentomaxillofac Radiol* 2004;33(1):6-11.

19. Hellen-Halme K, Rohlin M, Peterson A. Dental digital radiography. A survey of quality aspects. *Swed Dent J* 2005;29(2):81-7.

20. Wenzel A. A review of dentists' use of digital radiography and caries diagnosis with digital systems. *Dentomaxillofac Radiol* 2006;35(5):307-14.